

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01630	Calib. Date: 01/08/2020
Certification File No.:	01631	Calib. No.: 00047
Linearity File No.:	01632	Cert. Date: 01/08/2020
Solution File No.:	01647	Cert. No.: 00034
Sequential File No.:	01647	Lin. Date: 01/08/2020
		Lin. No.: 00033
		Soln. Date: 05/06/2020
		Soln. No.: 00266
		File Date: 05/06/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 0459

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:04D		
Control 1 EC	0.101%	09:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06D		
Control 2 EC	0.100%	09:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07D		
Control 3 EC	0.100%	09:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:09D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BAL

First Name: FRANK

MI: N

Signature: 

Badge No.: 467

Date: 05/06/2020