

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01724	Calib. Date:	09/21/2021
Certification File No.:	01725	Cert. Date:	09/21/2021
Linearity File No.:	01726	Lin. Date:	09/21/2021
Solution File No.:	01742	Soln. Date:	12/12/2021
Sequential File No.:	01742	File Date:	12/12/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	1466

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:34S		
Control 1 EC	0.098%	11:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35S		
Control 2 EC	0.098%	11:36S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37S		
Control 3 EC	0.098%	11:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LORETUCCI

First Name: KEVIN

MI: M

Signature: _____

Sgt Kevin #476

Badge No.: 476

Date: 12/12/2021