

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01770 Calib. Date: 07/12/2022 Calib. No.: 00057
Certification File No.: 01771 Cert. Date: 07/12/2022 Cert. No.: 00041
Linearity File No.: 01772 Lin. Date: 07/12/2022 Lin. No.: 00040
Solution File No.: 01796 Soln. Date: 11/28/2022 Soln. No.: 00311
Sequential File No.: 01796 File Date: 11/28/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1151

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:09S		
Control 1 EC	0.099%	12:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11S		
Control 2 EC	0.098%	12:11S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12S		
Control 3 EC	0.097%	12:12S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature:  # 85

Badge No.: 85

Date: 11/28/2022