

Calibrating Unit New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01798	Calib. Date:	12/13/2022
Certification File No.:	01799	Cert. No.:	00058
Linearity File No.:	01800	Cert. Date:	12/13/2022
Solution File No.:	01801	Lin. Date:	12/13/2022
Sequential File No.:	01801	Lin. No.:	00041
		Soln. Date:	12/13/2022
		Soln. No.:	00312
		File Date:	12/13/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1320

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:31S		
Control 1 EC	0.102%	15:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32S		
Control 2 EC	0.101%	15:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33S		
Control 3 EC	0.099%	15:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWA P2 - 215 (C8)

Changed By:
 Last Name: GUARCELLO First Name: COREY MI: R.
 Signature:  Badge No.: 7035
 Date: 12/13/2022