

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01770	Calib. Date:	07/12/2022
Certification File No.:	01771	Cert. Date:	07/12/2022
Linearity File No.:	01772	Lin. Date:	07/12/2022
Solution File No.:	01773	Soln. Date:	07/12/2022
Sequential File No.:	01773	File Date:	07/12/2022

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	09/01/2023
Solution Control Lot:	21360			Bottle No.:	1367

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:16D		
Control 1 EC	0.100%	16:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17D		
Control 2 EC	0.099%	16:17D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:18D		
Control 3 EC	0.099%	16:19D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWA P2-216 (8)

**Changed By:**

Last Name: GUARCELLO

First Name: COREY

MI: R.

Signature: 

Badge No.: 7035

Date: 07/12/2022