

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01751 Calib. Date: 02/24/2022 Calib. No.: 00056
Certification File No.: 01752 Cert. Date: 02/24/2022 Cert. No.: 00040
Linearity File No.: 01753 Lin. Date: 02/24/2022 Lin. No.: 00039
Solution File No.: 01769 Soln. Date: 07/04/2022 Soln. No.: 00305
Sequential File No.: 01769 File Date: 07/04/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1463

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:38D		
Control 1 EC	0.098%	08:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:39D		
Control 2 EC	0.097%	08:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41D		
Control 3 EC	0.096%	08:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: _____

Badge No.: 97

Date: 07/04/2022