

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01770	Calib. Date: 07/12/2022
Certification File No.:	01771	Calib. No.: 00057
Linearity File No.:	01772	Cert. Date: 07/12/2022
Solution File No.:	01777	Cert. No.: 00041
Sequential File No.:	01777	Lin. Date: 07/12/2022
		Lin. No.: 00040
		Soln. Date: 09/05/2022
		Soln. No.: 00308
		File Date: 09/05/2022

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %: 0.100%		Expires: 09/01/2023
Solution Control Lot: 21360		Bottle No.: 1386

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:17D		
Control 1 EC	0.099%	18:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:18D		
Control 2 EC	0.098%	18:19D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:19D		
Control 3 EC	0.097%	18:20D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: _____



#97

Badge No.: 97

Date: 09/05/2022