

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01798	Calib. Date: 12/13/2022
Certification File No.:	01799	Calib. No.: 00058
Linearity File No.:	01800	Cert. Date: 12/13/2022
Solution File No.:	01808	Cert. No.: 00042
Sequential File No.:	01808	Lin. Date: 12/13/2022
		Lin. No.: 00041
		Soln. Date: 01/09/2023
		Soln. No.: 00313
		File Date: 01/09/2023

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1316

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:34S		
Control 1 EC	0.101%	18:34S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:35S		
Control 2 EC	0.099%	18:36S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:36S		
Control 3 EC	0.099%	18:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: Edward J. McQuade # 97

Badge No.: 97

Date: 01/09/2023