

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01798	Calib. Date: 12/13/2022
Certification File No.:	01799	Calib. No.: 00058
Linearity File No.:	01800	Cert. Date: 12/13/2022
Solution File No.:	01815	Cert. No.: 00042
Sequential File No.:	01815	Lin. Date: 12/13/2022
		Lin. No.: 00041
		Soln. Date: 02/06/2023
		Soln. No.: 00314
		File Date: 02/06/2023

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %: 0.100%		Expires: 09/01/2023
Solution Control Lot: 21360		Bottle No.: 1254

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:44S		
Control 1 EC	0.101%	14:45S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46S		
Control 2 EC	0.100%	14:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47S		
Control 3 EC	0.100%	14:48S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: _____

Nicholas Larocca # 85

Badge No.: 85

Date: 02/06/2023