

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01798 Calib. Date: 12/13/2022 Calib. No.: 00058  
Certification File No.: 01799 Cert. Date: 12/13/2022 Cert. No.: 00042  
Linearity File No.: 01800 Lin. Date: 12/13/2022 Lin. No.: 00041  
Solution File No.: 01823 Soln. Date: 05/02/2023 Soln. No.: 00317  
Sequential File No.: 01823 File Date: 05/02/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364  
Control Solution %: 0.100% Expires: 09/01/2023  
Solution Control Lot: 21360 Bottle No.: 1303

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:18D		
Control 1 EC	0.101%	15:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 2 EC	0.100%	15:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:20D		
Control 3 EC	0.100%	15:21D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: Edward J. McQuade #97

Badge No.: 97

Date: 05/02/2023