

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01830	Calib. Date: 06/06/2023
Certification File No.:	01831	Cert. No.: 00059
Linearity File No.:	01832	Cert. Date: 06/06/2023
Solution File No.:	01841	Lin. Date: 06/06/2023
Sequential File No.:	01841	Soln. Date: 08/31/2023
		File Date: 08/31/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 0962

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:01D		
Control 1 EC	0.099%	13:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		
Control 2 EC	0.097%	13:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04D		
Control 3 EC	0.098%	13:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature:

Edward McQuade #97

Badge No.: 97

Date: 08/31/2023