

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00953	Calib. Date: 09/18/2012
Certification File No.:	00954	Calib. No.: 00028
Linearity File No.:	00955	Cert. Date: 09/18/2012
Solution File No.:	00960	Cert. No.: 00016
Sequential File No.:	00960	Lin. Date: 09/18/2012
		Lin. No.: 00016
		Soln. Date: 10/18/2012
		Soln. No.: 00126
		File Date: 10/18/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 0778

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:15D		
Control 1 EC	0.100%	11:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16D		
Control 2 EC	0.100%	11:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		
Control 3 EC	0.100%	11:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature: \_\_\_\_\_

*Sgt. T. Tarr*

Badge No.: 427

Date: 10/18/2012