

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01030	Calib. Date: 08/14/2013
Certification File No.:	01031	Calib. No.: 00030
Linearity File No.:	01032	Cert. Date: 08/14/2013
Solution File No.:	01048	Cert. No.: 00018
Sequential File No.:	01048	Lin. Date: 08/14/2013
		Lin. No.: 00018
		Soln. Date: 10/30/2013
		Soln. No.: 00144
		File Date: 10/30/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0114

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:21S		
Control 1 EC	0.099%	12:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23S		
Control 2 EC	0.098%	12:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24S		
Control 3 EC	0.099%	12:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 10/30/2013