

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00953	Calib. Date:	09/18/2012
Certification File No.:	00954	Cert. Date:	09/18/2012
Linearity File No.:	00955	Lin. Date:	09/18/2012
Solution File No.:	00970	Soln. Date:	11/15/2012
Sequential File No.:	00970	File Date:	11/15/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	12C100	Expires:	03/02/2014
		Bottle No.:	0418

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:09S		
Control 1 EC	0.101%	13:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10S		
Control 2 EC	0.100%	13:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:12S		
Control 3 EC	0.101%	13:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature: \_\_\_\_\_



Badge No.: 427

Date: 11/15/2012