

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01030	Calib. Date: 08/14/2013
Certification File No.:	01031	Calib. No.: 00030
Linearity File No.:	01032	Cert. Date: 08/14/2013
Solution File No.:	01055	Cert. No.: 00018
Sequential File No.:	01055	Lin. Date: 08/14/2013
		Lin. No.: 00018
		Soln. Date: 11/25/2013
		Soln. No.: 00145
		File Date: 11/25/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	12/26/2014
Solution Control Lot:	12L109			Bottle No.:	0124

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:36S		
Control 1 EC	0.100%	11:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37S		
Control 2 EC	0.099%	11:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:39S		
Control 3 EC	0.099%	11:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: MARYLOUISE

MI: -

Signature:

Sgt Marylouise Tarr #431

Badge No.: 431

Date: 11/25/2013