

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00841	Calib. Date:	10/26/2011
Certification File No.:	00842	Cert. Date:	10/26/2011
Linearity File No.:	00843	Lin. Date:	10/26/2011
Solution File No.:	00881	Soln. Date:	02/23/2012
Sequential File No.:	00881	File Date:	02/23/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	10H081	Expires:	08/04/2012
		Bottle No.:	0181

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:14S		
Control 1 EC	0.099%	13:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15S		
Control 2 EC	0.099%	13:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17S		
Control 3 EC	0.098%	13:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature: Sgt. P. Dan #427

Badge No.: 427

Date: 02/23/2012