

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01074	Calib. Date:	02/06/2014
Certification File No.:	01075	Cert. No.:	00031
Linearity File No.:	01076	Cert. Date:	02/06/2014
Solution File No.:	01081	Lin. Date:	02/06/2014
Sequential File No.:	01081	Soln. Date:	02/28/2014
		Lin. No.:	00019
		Soln. No.:	00150
		File Date:	02/28/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	13G115	Expires:	07/18/2015
		Bottle No.:	0995

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:16S		
Control 1 EC	0.101%	13:17S	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	13:17S	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	13:18S		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 02/28/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01074	Calib. Date: 02/06/2014
Certification File No.:	01075	Calib. No.: 00031
Linearity File No.:	01076	Cert. Date: 02/06/2014
Solution File No.:	01082	Lin. No.: 00019
Sequential File No.:	01082	Soln. Date: 02/28/2014
		Soln. No.: 00151
		File Date: 02/28/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	13G115	Expires: 07/18/2015
		Bottle No.: 0003

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:52S		
Control 1 EC	0.100%	14:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54S		
Control 2 EC	0.098%	14:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55S		
Control 3 EC	0.098%	14:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 02/28/2014