

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00953	Calib. Date: 09/18/2012
Certification File No.:	00954	Calib. No.: 00028
Linearity File No.:	00955	Cert. Date: 09/18/2012
Solution File No.:	00983	Cert. No.: 00016
Sequential File No.:	00983	Lin. Date: 09/18/2012
		Lin. No.: 00016
		Soln. Date: 02/04/2013
		Soln. No.: 00131
		File Date: 02/04/2013
Calibrating Unit:	1	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	12C100	Expires: 03/02/2014
		Bottle No.: 0785

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:58S		
Control 1 EC	0.102%	10:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59S		
Control 2 EC	0.101%	11:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00S		
Control 3 EC	0.100%	11:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCMANON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 02/04/2013