

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 00841 Calib. Date: 10/26/2011 Calib. No.: 00026
Certification File No.: 00842 Cert. Date: 10/26/2011 Cert. No.: 00014
Linearity File No.: 00843 Lin. Date: 10/26/2011 Lin. No.: 00014
Solution File No.: 00891 Soln. Date: 03/12/2012 Soln. No.: 00116
Sequential File No.: 00891 File Date: 03/12/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
Control Solution %: 0.100% Expires: 08/04/2012
Solution Control Lot: 10H081 Bottle No.: 0192

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:39S		
Control 1 EC	0.098%	12:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40S		
Control 2 EC	0.097%	12:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:42S		
Control 3 EC	0.097%	12:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____



Badge No.: 427

Date: 03/12/2012