

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: WEST WINDSOR POLICE
 Calibration File No.: 01074
 Certification File No.: 01075
 Linearity File No.: 01076
 Solution File No.: 01086
 Sequential File No.: 01086

Serial No.: ARWF-0382
 Calib. Date: 02/06/2014
 Cert. Date: 02/06/2014
 Lin. Date: 02/06/2014
 Soln. Date: 03/16/2014
 File Date: 03/16/2014
 Calib. No.: 00031
 Cert. No.: 00019
 Lin. No.: 00019
 Soln. No.: 00152

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 13G115

Model No.: CU-34
 Serial No.: DDWJ S3-0364
 Expires: 07/18/2015
 Bottle No.: 0994

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:25S		
Control 1 EC	0.099%	09:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27S		
Control 2 EC	0.096%	09:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28S		
Control 3 EC	0.097%	09:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 03/16/2014