

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01169	Calib. Date: 01/08/2015
Certification File No.:	01170	Calib. No.: 00033
Linearity File No.:	01171	Cert. Date: 01/08/2015
Solution File No.:	01201	Cert. No.: 00021
Sequential File No.:	01201	Lin. Date: 01/08/2015
		Lin. No.: 00021
		Soln. Date: 03/31/2015
		Soln. No.: 00173
		File Date: 03/31/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %:	0.100%		Expires: 08/22/2016
Solution Control Lot:	14H132		Bottle No.: 1255

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:17S		
Control 1 EC	0.100%	16:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19S		
Control 2 EC	0.097%	16:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20S		
Control 3 EC	0.099%	16:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 438

Date: 03/31/2015