

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01169	Calib. Date: 01/08/2015
Certification File No.:	01170	Calib. No.: 00033
Linearity File No.:	01171	Cert. Date: 01/08/2015
Solution File No.:	01216	Cert. No.: 00021
Sequential File No.:	01216	Lin. Date: 01/08/2015
		Lin. No.: 00021
		Soln. Date: 05/18/2015
		Soln. No.: 00175
		File Date: 05/18/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %:	0.100%		Expires: 08/22/2016
Solution Control Lot:	14H132		Bottle No.: 1259

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:26D		
Control 1 EC	0.099%	10:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:27D		
Control 2 EC	0.098%	10:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 3 EC	0.098%	10:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____

A. Latorre #68

Badge No.: 68

Date: 05/18/2015