

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00897	Calib. Date: 03/26/2012
Certification File No.:	00898	Calib. No.: 00027
Linearity File No.:	00899	Cert. Date: 03/26/2012
Solution File No.:	00922	Cert. No.: 00015
Sequential File No.:	00922	Lin. Date: 03/26/2012
		Lin. No.: 00015
		Soln. Date: 05/27/2012
		Soln. No.: 00120
		File Date: 05/27/2012

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	10H081	Expires: 08/04/2012
		Bottle No.: 0197

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:39D		
Control 1 EC	0.098%	19:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:40D		
Control 2 EC	0.097%	19:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	19:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:42D		
Control 3 EC	0.097%	19:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature: \_\_\_\_\_

*Sgt. Thomas*

Badge No.: 427

Date: 05/27/2012