

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00897	Calib. Date: 03/26/2012
Certification File No.:	00898	Calib. No.: 00027
Linearity File No.:	00899	Cert. Date: 03/26/2012
Solution File No.:	00926	Cert. No.: 00015
Sequential File No.:	00926	Lin. Date: 03/26/2012
		Lin. No.: 00015
		Soln. Date: 06/21/2012
		Soln. No.: 00121
		File Date: 06/21/2012

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 08/04/2012
Solution Control Lot:	10H081		Bottle No.: 0187

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:14D		
Control 1 EC	0.099%	11:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16D		
Control 2 EC	0.098%	11:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		
Control 3 EC	0.099%	11:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. T. Tarr

Badge No.: 427

Date: 06/21/2012