

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01579	Calib. Date: 03/07/2019
Certification File No.:	01580	Calib. No.: 00045
Linearity File No.:	01581	Cert. Date: 03/07/2019
Solution File No.:	01588	Cert. No.: 00032
Sequential File No.:	01588	Lin. Date: 03/07/2019
		Lin. No.: 00031
		Soln. Date: 04/10/2019
		Soln. No.: 00246
		File Date: 04/10/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0642

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:02D		
Control 1 EC	0.098%	16:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04D		
Control 2 EC	0.098%	16:05D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:05D		
Control 3 EC	0.098%	16:06D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BAL

First Name: FRANK

MI: N

Signature:



Badge No.: 467

Date: 04/10/2019