

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01316	Calib. Date: 03/28/2016
Certification File No.:	01317	Calib. No.: 00037
Linearity File No.:	01318	Cert. Date: 03/28/2016
Solution File No.:	01366	Cert. No.: 00024
Sequential File No.:	01366	Lin. Date: 03/28/2016
		Lin. No.: 00024
		Soln. Date: 08/17/2016
		Soln. No.: 00200
		File Date: 08/17/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 0869

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:13D		
Control 1 EC	0.098%	08:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14D		
Control 2 EC	0.099%	08:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		
Control 3 EC	0.100%	08:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 568

Date: 08/17/2016