

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01222 Calib. Date: 06/10/2015 Calib. No.: 00034
Certification File No.: 01223 Cert. Date: 06/10/2015 Cert. No.: 00022
Linearity File No.: 01224 Lin. Date: 06/10/2015 Lin. No.: 00022
Solution File No.: 01244 Soln. Date: 08/25/2015 Soln. No.: 00180
Sequential File No.: 01244 File Date: 08/25/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 04/13/2017
Solution Control Lot: 15D138 Bottle No.: 0004

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:34D		
Control 1 EC	0.097%	08:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35D		
Control 2 EC	0.097%	08:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37D		
Control 3 EC	0.097%	08:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: MARYLOUISE

MI: -

Signature: 

Badge No.: 431

Date: 08/25/2015