

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01222	Calib. Date: 06/10/2015
Certification File No.:	01223	Calib. No.: 00034
Linearity File No.:	01224	Cert. Date: 06/10/2015
Solution File No.:	01268	Cert. No.: 00022
Sequential File No.:	01268	Lin. Date: 06/10/2015
		Lin. No.: 00022
		Soln. Date: 11/06/2015
		Soln. No.: 00186
		File Date: 11/06/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 04/13/2017
Solution Control Lot:	15D138		Bottle No.: 0170

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:35S		
Control 1 EC	0.100%	21:36S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:36S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:37S		
Control 2 EC	0.098%	21:37S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:37S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:38S		
Control 3 EC	0.100%	21:39S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	21:39S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: _____



89

Badge No.: 89

Date: 11/06/2015

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01222	Calib. Date:	06/10/2015
Certification File No.:	01223	Cert. Date:	06/10/2015
Linearity File No.:	01224	Lin. Date:	06/10/2015
Solution File No.:	01267	Soln. Date:	11/06/2015
Sequential File No.:	01267	File Date:	11/06/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0364
Control Solution %:	0.100%			Expires:	04/13/2017
Solution Control Lot:	15D138			Bottle No.:	0169

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:43S		
Control 1 EC	-.---%	18:44S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	18:44S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____

Badge No.: 68

Date: 11/06/2015