

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01278	Calib. Date:	12/09/2015
Certification File No.:	01279	Cert. Date:	12/09/2015
Linearity File No.:	01280	Lin. Date:	12/09/2015
Solution File No.:	01289	Soln. Date:	12/29/2015
Sequential File No.:	01289	File Date:	12/29/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	0168

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:46S		
Control 1 EC	0.102%	13:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47S		
Control 2 EC	0.101%	13:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		
Control 3 EC	0.101%	13:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____

Badge No.: 68

Date: 12/29/2015