

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01222 Calib. Date: 06/10/2015 Calib. No.: 00034
Certification File No.: 01223 Cert. Date: 06/10/2015 Cert. No.: 00022
Linearity File No.: 01224 Lin. Date: 06/10/2015 Lin. No.: 00022
Solution File No.: 01232 Soln. Date: 07/07/2015 Soln. No.: 00178
Sequential File No.: 01232 File Date: 07/07/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 08/22/2016
Solution Control Lot: 14H132 Bottle No.: 1265

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:35D		
Control 1 EC	0.099%	16:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:36D		
Control 2 EC	0.098%	16:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:37D		
Control 3 EC	0.098%	16:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:39D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature:  #585

Badge No.: 585

Date: 07/07/2015