

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01604 Calib. Date: 07/29/2019 Calib. No.: 00046  
Certification File No.: 01605 Cert. Date: 07/29/2019 Cert. No.: 00033  
Linearity File No.: 01606 Lin. Date: 07/29/2019 Lin. No.: 00032  
Solution File No.: 01619 Soln. Date: 10/27/2019 Soln. No.: 00257  
Sequential File No.: 01619 File Date: 10/27/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364  
Control Solution %: 0.100% Expires: 09/26/2020  
Solution Control Lot: 18320 Bottle No.: 0156

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:50S		
Control 1 EC	0.099%	17:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51S		
Control 2 EC	0.099%	17:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:53S		
Control 3 EC	0.100%	17:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROWN

First Name: KYLE

M1: W

Signature: \_\_\_\_\_



Badge No.: 089

Date: 10/27/2019