

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01471	Calib. Date: 12/04/2017
Certification File No.:	01472	Calib. No.: 00042
Linearity File No.:	01473	Cert. Date: 12/04/2017
Solution File No.:	01496	Cert. No.: 00029
Sequential File No.:	01496	Lin. Date: 12/04/2017
		Lin. No.: 00028
		Soln. Date: 03/15/2018
		Soln. No.: 00229
		File Date: 03/15/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0661

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:31S		
Control 1 EC	0.098%	13:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33S		
Control 2 EC	0.098%	13:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	13:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34S		
Control 3 EC	0.097%	13:35S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: \_\_\_\_\_



Badge No.: 68

Date: 03/15/2018