

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE	Calib. No.:	00043
Calibration File No.:	01513	Calib. Date:	04/27/2018
Certification File No.:	01514	Cert. Date:	04/27/2018
Linearity File No.:	01515	Lin. Date:	04/27/2018
Solution File No.:	01538	Soln. Date:	08/09/2018
Sequential File No.:	01538	File Date:	08/09/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	0759

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:11D		
Control 1 EC	0.098%	22:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12D		
Control 2 EC	0.098%	22:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:13D		
Control 3 EC	0.098%	22:14D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 68

Date: 08/09/2018