

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01548 **Calib. Date:** 09/26/2018 **Calib. No.:** 00044
Certification File No.: 01549 **Cert. Date:** 09/26/2018 **Cert. No.:** 00031
Linearity File No.: 01550 **Lin. Date:** 09/26/2018 **Lin. No.:** 00030
Solution File No.: 01567 **Soln. Date:** 12/19/2018 **Soln. No.:** 00240
Sequential File No.: 01567 **File Date:** 12/19/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWJ S3-0364
Control Solution %: 0.100% **Expires:** 09/14/2019
Solution Control Lot: 17290 **Bottle No.:** 0202

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:26S		
Control 1 EC	0.098%	19:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:28S		
Control 2 EC	0.098%	19:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:29S		
Control 3 EC	0.098%	19:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: 

Badge No.: 89

Date: 12/19/2018