

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01413	Calib. Date: 02/22/2017
Certification File No.:	01414	Calib. No.: 00040
Linearity File No.:	01415	Cert. Date: 02/22/2017
Solution File No.:	01420	Cert. No.: 00027
Sequential File No.:	01420	Lin. Date: 02/22/2017
		Lin. No.: 00026
		Soln. Date: 03/23/2017
		Soln. No.: 00210
		File Date: 03/23/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	16330	Expires: 10/28/2018
		Bottle No.: 0977

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:15S		
Control 1 EC	0.100%	11:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16S		
Control 2 EC	0.100%	11:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17S		
Control 3 EC	0.100%	11:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOHR

First Name: DANNY

MI:

Signature: _____



Badge No.: 71

Date: 03/23/2017