

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>West Windsor Police 271 Clarksville Road West Windsor, NJ 08550</b>		2. Contact: <b>Sgt. Mohr</b>	4. Date: <b>11-16-2015</b>
		3. Phone Number: <b>609-799-2400</b>	5. County: <b>Mercer</b>
6. Alcotest Instrument Serial Number: <b>ARWF-0382</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Complaint received of instrument experiencing repeated control test failures.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

Last Known Sequential File #: 01275

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Michelle L. Goncalves #6040  
Name & Badge Number (Print)

*Sgt. Michelle L. Goncalves #6040* 11/16/15  
Signature Date

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: West Windsor P.D., West Windsor, NJ

Date Received: 11/20/15 Date given to service: 11/20/15

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  7110  8610  
 7510  DT5000  7410 Upper-half  9510  7410 Whole  
 Serial Number: AR WF-0382  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: October 2016 (59210)

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

\_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Fuel Cell = ARHJ-1714  
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 12/03/15



# Packing Slip

Customer no    Order No    Order date  
150059210    10820112    12/03/2015

Packing slip no.    Ship date  
80995025    12/03/2015

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	2 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>West Windsor Police  271 Clarksville Road  West Windsor, NJ 08550</b>	2. Contact: <b>Sgt. Mohr</b>	4. Date: <b>12-09-2015</b>
	3. Phone Number: <b>609-799-2400</b>	5. County: <b>Mercer</b>

6. Alcotest Instrument Serial Number: <b>ARWF-0382</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>
---	---	---

9. Reason for Service:  
**The above Alcotest was returned from outside evaluation and/or repair. A successful calibration was conducted utilizing the above Alcotest.**


10. Comments:  
**See attached packing slip and Drager Return and Repair Form.**

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Michelle L. Goncalves #6040  
Name & Badge Number (Print)

  
Signature

#6040 12/9/15  
Date