

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE	Calib. No.:	00047
Calibration File No.:	01630	Cert. Date:	01/08/2020
Certification File No.:	01631	Lin. Date:	01/08/2020
Linearity File No.:	01632	Soln. Date:	02/16/2020
Solution File No.:	01639	File Date:	02/16/2020
Sequential File No.:	01639		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	18320	Expires:	09/26/2020
		Bottle No.:	0261

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:00S		
Control 1 EC	0.101%	03:00S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:01S		
Control 2 EC	0.097%	03:02S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:03S		
Control 3 EC	0.096%	03:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 68

Date: 02/16/2020