

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01630	Calib. Date: 01/08/2020
Certification File No.:	01631	Calib. No.: 00047
Linearity File No.:	01632	Cert. Date: 01/08/2020
Solution File No.:	01643	Cert. No.: 00034
Sequential File No.:	01643	Lin. Date: 01/08/2020
		Lin. No.: 00033
		Soln. Date: 03/05/2020
		Soln. No.: 00264
		File Date: 03/05/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 0145

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:19S		
Control 1 EC	0.101%	00:20S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:21S		
Control 2 EC	0.097%	00:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:22S		
Control 3 EC	0.097%	00:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: \_\_\_\_\_



Badge No.: 68

Date: 03/05/2020