

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01649	Calib. Date:	06/10/2020
Certification File No.:	01650	Cert. No.:	00048
Linearity File No.:	01651	Lin. Date:	06/10/2020
Solution File No.:	01658	Cert. No.:	00035
Sequential File No.:	01658	Lin. No.:	00034
		Soln. No.:	00270
		File Date:	07/10/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	18320	Expires:	09/26/2020
		Bottle No.:	0801

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:38D		
Control 1 EC	0.102%	11:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:39D		
Control 2 EC	0.101%	11:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40D		
Control 3 EC	0.100%	11:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BAL

First Name: FRANK

MI: N

Signature: \_\_\_\_\_

Badge No.: 467

Date: 07/10/2020