Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARWF-0382

Location: WEST WINDSOR POLICE

Calibration File No.:	01649	Calib. Date:	06/10/2020	Calib. No.:	00048
Certification File No.:	01650	Cert. Date:	06/10/2020	Cert. No.:	00035
Linearity File No.:	01651	Lin. Date:	06/10/2020	Lin. No.:	00034
Solution File No.:	01673	Soln. Date:	11/16/2020	Soln. No.:	00275
Sequential File No.:	01673	File Date:	11/16/2020		

Calibrating Unit: WET — Model No.: CU-34 — Serial No.: DDWJ S3-0364

Control Solution %: 0.100% — Expires: 07/06/2022

Solution Control Lot: 20320 — Bottle No.: 1205

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:18S		
Control 1 EC	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19S		
Control 2 EC	0.099%	10:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21S		
Control 3 EC	0.099%	10:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE — First Name: EDWARD — MI: J

Signature: Det. #97 — Badge No.: 97

Date: 11/16/2020