

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
**Location:** WEST WINDSOR POLICE  
**Calibration File No.:** 01694 **Calib. Date:** 03/31/2021 **Calib. No.:** 00054  
**Certification File No.:** 01695 **Cert. Date:** 03/31/2021 **Cert. No.:** 00038  
**Linearity File No.:** 01696 **Lin. Date:** 03/31/2021 **Lin. No.:** 00037  
**Solution File No.:** 01701 **Soln. Date:** 04/28/2021 **Soln. No.:** 00286  
**Sequential File No.:** 01701 **File Date:** 04/28/2021

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDWJ S3-0364  
**Control Solution %:** 0.100% **Expires:** 07/06/2022  
**Solution Control Lot:** 20320 **Bottle No.:** 1202

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:21D		
Control 1 EC	0.100%	18:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:22D		
Control 2 EC	0.099%	18:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:23D		
Control 3 EC	0.099%	18:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: \_\_\_\_\_



Badge No.: 597

Date: 04/28/2021