

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01694 Calib. Date: 03/31/2021 Calib. No.: 00054  
Certification File No.: 01695 Cert. Date: 03/31/2021 Cert. No.: 00038  
Linearity File No.: 01696 Lin. Date: 03/31/2021 Lin. No.: 00037  
Solution File No.: 01707 Soln. Date: 06/24/2021 Soln. No.: 00288  
Sequential File No.: 01707 File Date: 06/24/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364  
Control Solution %: 0.100% Expires: 07/06/2022  
Solution Control Lot: 20320 Bottle No.: 1460

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:13D		
Control 1 EC	0.098%	21:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:14D		
Control 2 EC	0.098%	21:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15D		
Control 3 EC	0.098%	21:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature:  #97

Badge No.: 97

Date: 06/24/2021