Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARWF-0382

Location: WEST WINDSOR POLICE

Calibration File No.:	01694	Calib. Date:	03/31/2021	Calib. No.:	00054
Certification File No.:	01695	Cert. Date:	03/31/2021	Cert. No.:	00038
Linearity File No.:	01696	Lin. Date:	03/31/2021	Lin. No.:	00037
Solution File No.:	01714	Soln. Date:	07/18/2021	Soln. No.:	00289
Sequential File No.:	01714	File Date:	07/18/2021		

Calibrating Unit: WET — Model No.: CU-34 — Serial No.: DDWJ S3-0364

Control Solution %: 0.100% — Expires: 07/06/2022

Solution Control Lot: 20320 — Bottle No.: 1462

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:17D		
Control 1 EC	0.097%	11:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		
Control 2 EC	0.097%	11:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20D		
Control 3 EC	0.098%	11:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE — First Name: EDWARD — MI: J

Signature: PH. #97 — Badge No.: 597

Date: 07/18/2021