

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01724	Calib. Date: 09/21/2021
Certification File No.:	01725	Cert. No.: 00055
Linearity File No.:	01726	Cert. Date: 09/21/2021
Solution File No.:	01747	Lin. No.: 00039
Sequential File No.:	01747	Lin. Date: 09/21/2021
		Soln. No.: 00038
		Soln. Date: 01/09/2022
		Soln. No.: 00297
		File Date: 01/09/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	20320	Expires: 07/06/2022
		Bottle No.: 1467

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:43S		
Control 1 EC	0.098%	13:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45S		
Control 2 EC	0.098%	13:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 3 EC	0.098%	13:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: _____



Badge No.: 589

Date: 01/09/2022