

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01751 Calib. Date: 02/24/2022 Calib. No.: 00056
Certification File No.: 01752 Cert. Date: 02/24/2022 Cert. No.: 00040
Linearity File No.: 01753 Lin. Date: 02/24/2022 Lin. No.: 00039
Solution File No.: 01756 Soln. Date: 04/11/2022 Soln. No.: 00300
Sequential File No.: 01756 File Date: 04/11/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1131

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:51D		
Control 1 EC	0.099%	12:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52D		
Control 2 EC	0.098%	12:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 3 EC	0.098%	12:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LORETUCCI

First Name: KEVIN

MI: M

Signature: Sgt Kevin 476

Badge No.: 476

Date: 04/11/2022