

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE		
Calibration File No.: 01724	Calib. Date: 09/21/2021	Calib. No.: 00055
Certification File No.: 01725	Cert. Date: 09/21/2021	Cert. No.: 00039
Linearity File No.: 01726	Lin. Date: 09/21/2021	Lin. No.: 00038
Solution File No.: 01750	Soln. Date: 02/06/2022	Soln. No.: 00298
Sequential File No.: 01750	File Date: 02/06/2022	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %: 0.100%		Expires: 07/06/2022
Solution Control Lot: 20320		Bottle No.: 1000

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:46S		
Control 1 EC	0.101%	07:46S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47S		
Control 2 EC	0.099%	07:47S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:48S		
Control 3 EC	0.099%	07:49S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: _____



89

Badge No.: 89

Date: 02/06/2022