Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE

Calibration File No.:	00897	Calib. Date:	03/26/2012	Calib. No.:	00027
Certification File No.:	00898	Cert. Date:	03/26/2012	Cert. No.:	00015
Linearity File No.:	00899	Lin. Date:	03/26/2012	Lin. No.:	00015
Solution File No.:	00911	Soln. Date:	04/30/2012	Soln. No.:	00119
Sequential File No.:	00911	File Date:	04/30/2012		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0364
Control Solution %:	0.100%			Expires:	08/04/2012
Solution Control Lot:	10H081			Bottle No.:	0193

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:29D		
Control 1 EC	0.099%	14:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31D		
Control 2 EC	0.098%	14:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32D		
Control 3 EC	0.098%	14:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: TARR First Name: THOMAS MI: S

Signature: Sgt. T. Tarr Badge No.: 427
Date: 04/30/2012