

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01118	Calib. Date: 07/09/2014
Certification File No.:	01119	Calib. No.: 00032
Linearity File No.:	01120	Cert. Date: 07/09/2014
Solution File No.:	01125	Cert. No.: 00020
Sequential File No.:	01125	Lin. Date: 07/09/2014
		Lin. No.: 00020
		Soln. Date: 08/13/2014
		Soln. No.: 00159
		File Date: 08/13/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	13F114	Expires: 06/17/2015
		Bottle No.: 0416

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:37D		
Control 1 EC	0.100%	10:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39D		
Control 2 EC	0.098%	10:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40D		
Control 3 EC	0.099%	10:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROWN

First Name: KYLE

MI: B

Signature: \_\_\_\_\_



Badge No.: 089

Date: 08/13/2014