

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01278 Calib. Date: 12/09/2015 Calib. No.: 00036
Certification File No.: 01279 Cert. Date: 12/09/2015 Cert. No.: 00023
Linearity File No.: 01280 Lin. Date: 12/09/2015 Lin. No.: 00023
Solution File No.: 01309 Soln. Date: 03/15/2016 Soln. No.: 00193
Sequential File No.: 01309 File Date: 03/15/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 04/13/2017
Solution Control Lot: 15D138 Bottle No.: 0165

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:48S		
Control 1 EC	0.101%	07:48S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.103%	07:48S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:49S		
Control 2 EC	0.101%	07:50S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	07:50S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:50S		
Control 3 EC	0.101%	07:51S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.103%	07:51S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature:  #585

Badge No.: 585

Date: 03/15/2016