Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE				
Calibration File No.:	01604	Calib. Date:	07/29/2019	Calib. No.:	00046
Certification File No.:	01605	Cert. Date:	07/29/2019	Cert. No.:	00033
Linearity File No.:	01606	Lin. Date:	07/29/2019	Lin. No.:	00032
Solution File No.:	01607	Soln. Date:	07/29/2019	Soln. No.:	00253
Sequential File No.:	01607	File Date:	07/29/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0364
Control Solution %:	0.100%			Expires:	09/26/2020
Solution Control Lot:	18320			Bottle No.:	0219

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:50D		
Control 1 EC	0.100%	11:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51D		
Control 2 EC	0.100%	11:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		
Control 3 EC	0.100%	11:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWAP2-215 BMG

Changed By:

Last Name: GAMBONE First Name: BRIAN MI: M

Signature: TPR. I B[illegible] # 7029 Badge No.: 7029 Date: 07/29/2019